

MANUFACTURER DECLARATION

FUELSTAT® Kit Contents for Import Purposes

PRODUCT DETAILS			
Common name/ Product type	Brand name	Description of individual test contained in sealed foil pouch	Package details (total size and size per unit)
Test kits using Lateral Flow Devices (LFD) for detection of microbial contamination in kerosene and diesel fuels. <i>Note 1</i>	FUELSTAT <i>resinae</i> Plus and FUELSTAT Diesel Plus	Tests comprise 6 LFDs each containing between 0.1 and 1.0 ug of antibody and between 0.1 and 0.6 ug of antigenic material both immobilised on a membrane (Note 1). The antibodies are raised against antigenic material which comprises a mixture of metabolites and cell wall material produced from one of: <i>Hormoconis resinae</i> <i>Psuedomonas putida</i> <i>Yarrowia lipolytica</i> Antigenic material is gamma irradiation sterilized in a commercial facility before use on the LFDs. <ul style="list-style-type: none"> • Pouch also contains a non- biological extraction fluid and a silica gel descant bag. <i>Note 2</i> • Extraction fluid is contained in sample bottle. <i>Note 3</i> 	8 test kits per box 1 box size: 360mm x 185mm x 225 mm (L x W x H) (Weight 1.35 kgs)

NOTE 1:

LFDs are plastic housings containing:-

Cellulose/ nitrocellulose (CAS No: 9004-70-0) membrane filter pads, absorbent paper pad, paper sample pad, backing strip, colloidal gold, polyester conjugate pad. Quantities of nitrocellulose are not classified as hazardous for transport (UN 3270, IATA ID A122 refers).

Colloidal gold comprises less than 1 mg in total per complete test paddle; i.e. per 6 LFDs. This equates to <0.002% of the test paddle and <0.0008% of the complete pouched test paddle and sample bottle.

NOTE 2:

Extraction fluid-15 mLs of buffer salt solution, surfactant and blue food colouring preserved using ProClin 950. The active ingredient of ProClin 950 is 2-Methyl-4-isothiazolin-3-one (MIT) (CAS-No: 2682-20-4) at 9.5-9.9% solution. **See Doc 302 for further details:**

<https://conidia.com/resources/fuelstat-literature/#documents>

Extraction fluid	
chemical	[%]
NaCl	0.8
KCl	0.02
KH ₂ PO ₄	0.02
Na ₂ HPO ₄ .7H ₂ O	0.547
Tween 20	0.1
ProClin 950	0.06
Ice Blue Dye	0.0025
Distilled Water	98.4505

NOTE 3:

Sample bottles and lids manufactured from Polypropylene random copolymers.

Revision Date: 26th September 2022

Version number: 4.0

Revision Comments: Updated details on ProClin 950

The information provided in this declaration is correct to the best of our knowledge, information and belief at the date of its publication. The information given is designed only as a guidance for safe handling, use, processing, storage, transport, disposal and release and is not to be considered a warranty or quality specification. The information relates only to the specific material designated and may not be valid for such material used in combination with any other materials or in any process, unless specified in the text.

Who we are:

FUELSTAT® fuel tests are developed, manufactured and marketed by Conidia Bioscience Limited. Based in UK, Conidia Bioscience Limited was founded in early 2000's by experts in immunoassay techniques and holds the internationally patented intellectual property for FUELSTAT®.

FUELSTAT® Result is hosted by Conidia Bioscience Limited and its service partners on secure servers and does not share any data with third parties.

Where to find us:

FUELSTAT® is distributed globally by a network of specialist distributors covering the major sectors. Contact info@conidia.com who will arrange for a distributor to support you.

HEADQUARTERS & GLOBAL SALES OFFICE

Conidia Bioscience Ltd
Bakeham Lane, Egham,
Surrey, TW20 9TY, UK
+44 (0)1491 829102
info@conidia.com

**Conidia
Bioscience**
WWW.CONIDIA.COM

US SALES OFFICE

Conidia Bioscience Inc
15 Briarwood Ln, Dover,
NH, 03820, USA
+1 407 616 0315
info@conidia.com